

# United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/509,457	05/02/2005	Sabine Mollus	DE020086US	2736
24737 PHILIPS INTE	7590 01/04/2009 ELLECTUAL PROPER	EXAMINER		
P.O. BOX 300	1	RAMIREZ, JOHN FERNANDO		
BRIARCLIFF MANOR, NY 10510			ART UNIT	PAPER NUMBER
			3737	
			MAIL DATE	DELIVERY MODE
			01/04/2008	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

		η,				
	Application No.	Applicant(s)				
	10/509,457	MOLLUS ET AL.				
Office Action Summary	Examiner	Art Unit				
	John F. Ramirez	3737				
The MAILING DATE of this communication app Period for Reply	pears on the cover sheet with th	ne correspondence address				
A SHORTENED STATUTORY PERIOD FOR REPLY WHICHEVER IS LONGER, FROM THE MAILING D/ Extensions of time may be available under the provisions of 37 CFR 1.1: after SIX (6) MONTHS from the mailing date of this communication.  If NO period for reply is specified above, the maximum statutory period v Failure to reply within the set or extended period for reply will, by statute Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	ATE OF THIS COMMUNICAT 36(a). In no event, however, may a reply build apply and will expire SIX (6) MONTHS 1, cause the application to become ABAND	ION. se timely filed from the mailing date of this communication. DNED (35 U.S.C. § 133).				
Status						
1) Responsive to communication(s) filed on 23 O	<u>ctober 2007</u> .					
Pa)⊠ This action is <b>FINAL</b> . 2b)□ This action is non-final.						
3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is						
closed in accordance with the practice under E	Ex parte Quayle, 1935 C.D. 11	, 453 O.G. 213.				
Disposition of Claims						
4) Claim(s) is/are pending in the application	n.					
4a) Of the above claim(s) is/are withdrawn from consideration.						
5) Claim(s) is/are allowed.						
6)⊠ Claim(s) <u>1-10</u> is/are rejected.						
7) Claim(s) is/are objected to.						
8) Claim(s) are subject to restriction and/o	r election requirement.					
Application Papers						
9) The specification is objected to by the Examine	r.					
10) ☐ The drawing(s) filed on is/are: a) ☐ acc	epted or b)□ objected to by th	ne Examiner.				
Applicant may not request that any objection to the		, ,				
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).						
11) The oath or declaration is objected to by the Ex	caminer. Note the attached Off	ice Action or form PTQ-152.				
Priority under 35 U.S.C. § 119						
12) Acknowledgment is made of a claim for foreign	priority under 35 U.S.C. § 119	θ(a)-(d) or (f).				
a) ☐ All b) ☐ Some * c) ☐ None of:  1. ☐ Certified copies of the priority documents have been received.						
Certified copies of the priority documents have been received in Application No						
3. Copies of the certified copies of the priority documents have been received in this National Stage						
application from the International Bureau	application from the International Bureau (PCT Rule 17.2(a)).					
* See the attached detailed Office action for a list of the certified copies not received.						
Attachment(s)						
Notice of References Cited (PTO-892)  Notice of Draftsperson's Patent Drawing Review (PTO-948)	4) Interview Summ Paper No(s)/Ma					
B) Information Disclosure Statement(s) (PTO/SB/08)	5) 🔲 Notice of Inform					
Paper No(s)/Mail Date	6)  Other:					

10/509,457 Art Unit: 3737

#### **DETAILED ACTION**

## Response to Amendment

After a careful review of applicant's remarks dated on 10/23/07 and 03/15/06 and 10/05/06 with respect to claims 1-10 have been fully considered and are not persuasive.

With respect to claims 1-10, allowability has been withdrawn and claims have been rejected in view of prior art, non-statutory subject matter and indefiniteness.

In relation to the section 102(b) rejection that anticipated claims 1, 4-6, 8-10 using Ohe, applicant alleges that the Ohe disclosure is limited in that "only one reference instant is used for the cardiac cycle". However, the examiner notes that simply duplicating the number of references instant to avoid undesirable and disturbing artifacts in the differential sequence should not be given patentable weight since such process is a mere duplication of the system as disclosed by Ohe (US 4,729,379). Duplicating the components of a prior art device is a design consideration within the skill of the art. In re Harza, 274 F.2d 669, 124 USPQ 378 (CCPA 1960). Therefore previous rejections with respect to these claims still stand.

# Claim Rejections - 35 USC § 112

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 1-10 are indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

Application/Control Number:

10/509,457 Art Unit: 3737

The terms "determining" and "determine" in claims 1-10 are a relative terms which renders the claims indefinite. The terms "determining" and "determine" are not defined by the claims, the specification does not provide a standard for ascertaining the requisite degree, and one of ordinary skill in the art would not be reasonably apprised of the scope of the invention.

### Claim Rejections - 35 USC § 101

35 U.S.C. 101 reads as follows:

Whoever invents or discovers any new and useful process, machine, manufacture, or composition of matter, or any new and useful improvement thereof, may obtain a patent therefor, subject to the conditions and requirements of this title.

Claim 10 is rejected under 35 U.S.C. 101 because the claimed invention is directed to non-statutory subject matter. The claim is directed to a computer program, which constitutes non-statutory subject matter.

#### Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:


A person shall be entitled to a patent unless -

Claims 1, 4-6, 8-10 are rejected under 35 U.S.C. 102(b) as being anticipated by Ohe (US 4,729,379).


Application/Control Number:

10/509,457 Art Unit: 3737

Page 4



-I CARDIAC BEAT -



With respect to claim 1, Ohe shows in Figures 4 and 6, a method of determining a corresponding image for a reference image from an image sequence of a moving object by means of a first and a second motion signal, in which the first and the second

10/509.457

Art Unit: 3737

motion signal represent the respective variation in time of the states of motion of a first motion and a second motion of the object, the image sequence represents the first motion of the object as a sequence of images of states of motion, the reference image represents a state of motion from the second object motion and is acquired at a reference instant during the second motion of the object including the following steps: a) determining a similarity function by way of a similarity comparison of the first and the second motion signal, b) determining a correspondence instant in the first motion signal by means of the similarity function, the correspondence instant corresponding to the acquisition instant of the reference image from the second motion signal, c) determining, using the first motion signal, that image of the image sequence whose acquisition constant corresponds at least approximately to the correspondence instant.

With respect to claims 4-6, Ohe shows in figures 1, 2A and 2B a method characterized in that the first and the second motion signal form an electrocardiopaphic signal and that the images of the image sequence and the reference image represent states of motion of a human or animal heart, characterized in that the blood vessels of the heart are filled at least partly with a contrast medium either in images of the image sequence or in the reference image, characterized in that the image sequence forms an X-ray image sequence and/or the reference image forms an X-ray image (col. 2, lines 22-57).

With respect to claims 8-10, Ohe shows in Figure 3 a system which includes a data processing unit for determining a corresponding image of a moving object for a reference image from an image sequence by means of a first motion signal (50,

Figure 3), the data processing unit being arranged to carry out a method as claimed in claim 1, an examination apparatus which includes an X-ray image detector and means for the detection of electrocardiographic signals, which apparatus includes a system as claimed in claim 1 (90, Figure 3), a computer program or computer program product which is arranged to cooperate with a data processing unit in such a manner that the data processing unit is capable of carrying out a method as claimed in claim 1 (50, Figure 3).

## Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claims 2, 3 and 7 are rejected under 35 U.S.C. 103(a) as being unpatentable over Ohe in view of Vullings (*Automated ECG segmentation with Dynamic Time Warping*) in further view of Beier (*Advanced Subtraction Angiography: Mask selection and Image Registration*) and in further view of Urbano et al. (US 6,228,030).

In reference to claims 2, 3 and 7, Ohe teaches all the limitations of the claimed subject matter except for mentioning specifically the steps of characterizing in that the similarity function is obtained by means of the so-called dynamic time warping method, characterizing in that an interpolation image is formed from the corresponding image and a further image from the image sequence, which interpolation image represents at

10/509,457 Art Unit: 3737

least substantially the state of motion of the object at the correspondence instant, and additionally, characterizing in that the image sequence and/or the reference image forms an ultrasound image.

However, the steps of 1) characterizing in that the similarity function is obtained by means of the so-called dynamic time warping method, 2) characterizing in that an interpolation image is formed from the corresponding image and a further image from the image sequence, which interpolation image represents at least substantially the state of motion of the object at the correspondence instant, and 3) characterizing in that the image sequence and/or the reference image forms an ultrasound image are considered conventional in the art as evidenced by the teachings of Vullings (Automated ECG segmentation with Dynamic Time Warping), Beier (Advanced Subtraction Angiography: Mask selection and Image Registration) and Urbano et al. (US 6,228,030).

The Vullings publication teaches the step of characterizing in that the similarity function is obtained by means of the so-called dynamic time warping method. Moreover, the Beier publication teaches the step of characterizing in that an interpolation image is formed from the corresponding image and a further image from the image sequence, which interpolation image represents at least substantially the state of motion of the object at the correspondence instant, and Furthermore, the Urbano et al., patent teaches the step of characterizing in that the image sequence and/or the reference image forms an ultrasound image.

Application/Control Number:

10/509,457 Art Unit: 3737

Based on the above observations, for a person of ordinary skill in the art, modifying the method disclosed by Ohe, with the above discussed enhancements would have been considered obvious because such modifications would have a great effect upon image quality.

#### Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to John F. Ramirez whose telephone number is (571) 272-8685. The examiner can normally be reached on (Mon-Fri) 7:00 - 3:30 p.m.

10/509,457 Art Unit: 3737

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Brian L. Casler can be reached on (571) 272-4956. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

JFR

Blel